

ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

System: Ceilometer

Site: Darwin

OLD Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Date UTC	Time Start UTC	Time End UTC
Laptop IRM 600	78-FY330	I-Van	n/a			8 December 2004	0110	0600

NEW Configuration

Part Number	Serial Number	Location	Configuration Version	Change Item 14	Change Item 15	Copy Config. Logger to PC	Save to EPROM(E)	Update Version #
Laptop IRM 600	78-FZG25	I-Van	n/a					

Replacement completed by: Troy Culgan

NOTES:

Old Part remains on site as spare: ☐

Old part packed for shipment to: _____

Other old part disposition (explain): _____

Other NOTES: Laptop replaced due to failed monitor, delay in bringing system back on line due to password for ftp
Not being known by technician, problem resolved by installing the unserviceable laptops hard drive in the replacement laptop.